


Edition D-P002-61-70-M0-1000, August 2023

Multi Joystick

Multi Joystick Omni (P002-61) - Multi Joystick R-net (P002-62) - Multi Joystick LiNX (P002-64)



Contents

Other languages	2
About this manual	3
User manual	3
mo-vis bv	3
User manual	5
Warning labels	5
Support, scrapping & recycling	8
Limited liability	9
Intended Use	9
Features	10
Operation	13
Parts and accessories	14
Qualifications	17
First time use	18
Cleaning	18
Warnings and limitations	19
Connection options	19
Combinations	21
Dimensions	21

Other languages

CS: Tito příručku naleznete v češtině na našich webových stránkách.
DA: Du kan finde denne manual på dansk på vores websted.
DE: Auf unserer Website finden Sie diese Anleitung in Deutsch.
EN: You can find this manual in English on our website.
ES: Este manual está disponible en español en nuestro sitio web.
FR: Vous pouvez trouver ce manuel en français sur notre site Web.
IT: nostro sito web è disponible il presente manuale in italiano.
NO: Du finner denne håndboken på norsk på nettstedet vårt.
NL: De Nederlandse handleiding kunt u op onze website vinden.
PT: Pode encontrar este manual em português na nossa página web.
SL: Slovenski priročnik je na voljo na naši spletni strani.
SV: Du hittar den här handboken på svenska på vår webbplats.

You can find the latest version of all our manuals on the website.

About this manual

User manual

Thank you for choosing a mo-vis product!

This manual contains **useful and important information** about your device. **Please read it carefully before use and store safely for future reference**.

Our team (or your authorized dealer) will be happy to answer your questions.

mo-vis bv

Biebuyckstraat 15D . 9850 Deinze . Belgium http://www.mo-vis.com . contact@mo-vis.com . +32 9 335 28

	60
CH REP	CH-REP: SKS Rehab AG, Im Wyden, 8762 Schwanden GL, Switzerland.
UK CA	UK Responsible Person: QServe Group UK, Ltd., 49 Greek Street, London, WID 4EG, UK.

User manual

Warning labels

Please read this manual, the safety instructions and warning texts carefully, in order to reduce the risks associated to the device. Our products are safe under normal and reasonably foreseeable operating conditions.

NOTE: This symbol indicates general notes and information.

CAUTION: This symbol indicates caution for a hazardous situation that, if not avoided, could result in minor or moderate injury.

WARNING: This symbol indicates a warning for a hazardous situation that, if not avoided, could result in death or serious injury.

Other labels:



Catalogue number: indicates the manufacturer's catalogue number so that the medical device can be

	identified.
LOT	Batch code: indicates the manufacturer's batch code so that the batch or lot can be identified.
MD	Medical device: indicates that the item is a medical device.
	Date of manufacture: indicates the date when the medical device was manufactured.
SN	Serial number: indicates the manufacturer's serial number so that a specific medical device can be identified.
	Consult instructions for use or con- sult electronic instructions for use: indicates the need for the user to consult the instructions for use.
	Keep dry: indicates a medical device that needs to be protected from moisture.
	Do not use if package is damaged and consult instructions for use:

	indicates that a medical device should not be used if the package has been damaged or opened and that the user should consult the instructions for use for additional information.
CE	CE label: indicates that the manufac- turer or importer affirms the good's conformity with European health, safety, and environmental protec- tion standards.
X	WEEE: indicates that the product should not be discarded as unsorted waste but must be sent to separate collection facilities for recovery and recycling.
	Manufacturer: indicates the medical device manufacturer.

Support, scrapping & recycling

Technical support

TROUBLE: In case of technical problems:

- Contact your dealer
- If your dealer is not available or unknown, please contact mo-vis: contact@mo-vis.com or +32 9 335 28 60.


Always state the product code and device serial number when contacting us. This ensures you are provided with the correct information.

Spare parts and accessories

Contact mo-vis or your dealer for more information about spare parts and accessories.


Scrapping & recycling

The expected lifetime of this device is 5 years.



Limited liability

mo-vis accepts **no liability** for personal injury or damage to property that may arise from the failure of the user or other persons to follow the recommendations, warnings and instructions in this manual.



NOTE: All mo-vis user manuals can be found at http://www.mo-vis.com where they can be consulted in PDF format.

NOTE: In case any serious incident occurs in relation to this device, this should be reported immediately to mo-vis and the competent authority of the Member State in which the user is established.

Intended Use

The Multi Joystick:

• is a joystick module which can be connected directly to the wheelchair electronics to control the wheelchair functions

- is a small proportional joystick, which requires significant less power and movement compared to a normal joystick
- can easily be controlled by any body part, including finger, chin, lip... thanks to its flexibility
- is recommended for people with tetraplegia, neuromuscular diseases (ALS, MS...) and limited muscular force

Features

The Multi Joystick is:

- Fully proportional, with reduced throw (8 mm 0.31 in) and force (ca. 50 gr 1.76 oz)
- Built in a compact and discrete housing, with two 3.5 mm jacks for switch input



 has 2 switches that can be linked to the separate interface: (1) on/off power wheelchair - (2) mode of the power wheelchair.



- is fully adjustable to individual possibilities and needs with well-thoughtout electronics (USB-C connection).
- is compatible with multiple types of wheelchair electronics.
- has a LED light on the interface which will start to flash green in case of an error or orange in case of active tilt (see *Installation manual*).

Operation

The movements of the control are translated into according movements of the wheelchair, e.g. driving or menu navigation.

Common practice to navigate the wheelchair with the joystick is as follows:

- **Direction**: point the control into the direction you want the wheelchair to move. The wheelchair then moves in that direction.
- **Speed**: the further you move the control from the default (center) position, the faster the wheelchair moves.
- **Stop**: whenever you release the control, the control moves back to the default (center) position and the wheelchair stops.

CAUTION: Avoid hitting obstacles during driving.

WARNING: When the LED light flashes and/or after every incident with the wheelchair or the mo-vis device, contact your dealer immediately to perform a functional test.

LED status

The illumination of the LED on the joystick interface indicates the operational status of the joystick.

JOYSTICK STATUS	COLOUR	LED TIMING	COUNT	STATE DESCRIPTION
Configur- ing	Orange	Fast	Not applicable	Wheelchair is configur- ing
Powercy- cle	Orange	Heartbeat		Wheelchair waits for power cycle
Focus	Green	Always on		System is in focus
Out of focus	Green	Heartbeat		System is out of focus
Out of neutral	Green	Medium		Joystick is out of neu- tral
Error	Red	Errors	See	Something went wrong

Parts and accessories

Package

The Multi Joystick package consists of:

EXAMPLE	P002-61 - MULTI JOYSTICK OMNI	P002-62 - MULTI JOYSTICK R-NET	P002-63 - MULTI JOYSTICK LINX
8	P002-60 - Multi Joystick Sensor		
	M002-05 - Multi Joystick ball		
	M002-18 - Multi Joystick Mounting set		
See Installation Manual	SUB D9 cable and Joystick Omni Interface	R-net cable and Joystick R-net Interface	LiNX cable and Joystick LiNX Interface

EXAMPLE	P002-61 - MULTI	P002-62 - MULTI	P002-63 - MULTI
	JOYSTICK OMNI	JOYSTICK R-NET	JOYSTICK LINX

D-P002-61-70-MX Multi Joystick User Manual

D-P002-61-70-M1 Multi Joystick Installation Manual

Optional

A second topper is optionally available: M002-53 Multi Joystick Cup.

Unit detail

The connections of the Multi Joystick unit have protective covers to avoid the intrusion of dust and moisture during usage.

CAUTION: Before inserting a connector, remove the protective cover.

CAUTION: Always put the protective covers in, or keep them, in case the connections are not used.



Spare parts

Spare parts and accessories must be ordered by the dealer at mo-vis by. Always mention the product code of the desired item.

Qualifications

CAUTION: Only a qualified service engineer may install the device.

CAUTION: An incorrect programming of the wheelchair or device electronics, may cause damage to the devices or injury to the user.

First time use



CAUTION: A qualified service engineer should assist during first time use of the device and make sure that all functions are clear to the user. including what to do in case of problems or doubt.

Cleaning

WARNING: As dust and dirt could lead to reduced functionality, all parts of the device should be cleaned on a regular basis (monthly) or whenever needed

- Gently remove dust and dirt with a damp cloth. ٠
- Use only non-aggressive disinfectant cleaning agents.

CAUTION: Do not immerse in water or use excessive amounts of liquid.

CAUTION: Do not apply additional lubricants to the moving parts.

The handles of the joystick can be removed to clean them.

- You may shortly immerse the handle in warm water with a gentle cleaning agent. Let the handle dry before putting it back on.
- If the handle is damaged, worn out or remains dirty, please contact your • authorized dealer or mo-vis to replace it.

Warnings and limitations

NOTE: This product complies with the limit values for Electromagnetic Compatibility (EMC) with respect to power wheelchairs, as set out in the harmonized standards for the EU in the Medical Device Regulation, No. 2017/745.



WARNING: Do not cover or block the device to avoid uncontrolled behavior of the device or the wheelchair.

CAUTION: If an output inhibit is programmed, then it should be checked monthly if this is still operational. Please see the Installation Manual for more information.

Connection options

The connection interface depends on the control version. Please see the Installation Manual for more information.



You can connect 1 or 2 switches to the unit:

- Red = power on/off
- Yellow = mode


WARNING: The on/off switch must be available to the user at all times. This allows to instantly stop the wheelchair in case of problems or an emergency. WARNING: Always power off the device when the user is no longer in the wheelchair to avoid unauthorized access or accidental movements.

Combinations


There are different versions of this device available for combination with different wheelchair electronics. Please refer to the *Installation Manual* of this device for more information.

Dimensions

Bottom view:



Side view:



PART	DIMENSIONS (METRIC)	DIMENSIONS (IMPERIAL)
Joystick	65 mm x 31 mm x 31	2.55 in x 1.22 in x 1.22

PART	DIMENSIONS (METRIC)	DIMENSIONS (IMPERIAL)
	mm (HxWxD)	in
Ball	26 mm x 30 mm x 30 mm	1.02 in x 1.18 in x 1.18 in
Cup	22 mm x 24 mm x 24 mm	0.87 in x 0.94 in x 0.94 in



mo-vis bv . Biebuyckstraat 15D . 9850 Deinze - Belgium www.mo-vis.com . contact@mo-vis.com . +32 9 335 28 60

Go to our website for more information on our products or share your experience with us via email.

