

# User manual

Edition D-P005-50-70-M0-0802, March 2026

## Twister

Twister Basic - Twister Button - Twister on Bended Tube - Satellite Twister  
- Twister Gooseneck



**mo-vis**  
moving forward together

# Contents

<b>Other languages .....</b>	<b>2</b>
<b>About this manual .....</b>	<b>3</b>
User manual .....	3
mo-vis bv .....	3
<b>User manual .....</b>	<b>4</b>
Warning labels .....	4
Warranty .....	5
Important information .....	6
Support, scrapping & recycling .....	6
Intended use .....	7
Features .....	7
Operation .....	7
Parts and accessories .....	8
Cleaning .....	12
Limited liability .....	13

## Other languages

CS: Tito příručku naleznete v češtině na našich webových stránkách.

DA: Du kan finde denne manual på dansk på vores websted.

DE: Auf unserer Website finden Sie diese Anleitung in Deutsch.

EN: You can find this manual in English on our website.

ES: Este manual está disponible en español en nuestro sitio web.

FR: Vous pouvez trouver ce manuel en français sur notre site Web.

IT: nostro sito web è disponibile il presente manuale in italiano.

NO: Du finner denne håndboken på norsk på nettstedet vårt.

NL: De Nederlandse handleiding kunt u op onze website vinden.

PT: Pode encontrar este manual em português na nossa página web.

SV: Du hittar den här handboken på svenska på vår webbplats.

**You can find the latest version of all our manuals on the website.**

# About this manual

## User manual

Thank you for choosing a mo-vis product!



This manual contains **useful and important information** about your device. **Please read it carefully before use and store safely for future reference.**

Our team (or your authorized dealer) will be happy to answer your questions.

## mo-vis bv

	Biebuyckstraat 15D . 9850 Deinze . Belgium <a href="http://www.mo-vis.com">http://www.mo-vis.com</a> <a href="mailto:contact@mo-vis.com">contact@mo-vis.com</a> +32 9 335 28 60		
<table border="1"><tr><td>CH</td><td>REP</td></tr></table>	CH	REP	CH-REP: SKS Rehab AG, Im Wyden, 8762 Schwanden GL, Switzerland.
CH	REP		
	UK Responsible Person: QServe Group UK, Ltd., 49 Greek Street, London, W1D 4EG, UK.		

# User manual

## Warning labels

*Please read this manual, the safety instructions and warning texts carefully, in order to reduce the risks associated to the device. Our products are safe under normal and reasonably foreseeable operating conditions.*



**NOTE:** This symbol indicates general notes and information.







**CAUTION:** This symbol indicates caution for a hazardous situation that, if not avoided, could result in minor or moderate injury.



**WARNING:** This symbol indicates a warning for a hazardous situation that, if not avoided, could result in death or serious injury.

Other labels:

	Catalogue number: indicates the manufacturer's catalogue number so that the medical device can be identified.
	Batch code: indicates the manufacturer's batch code so that the batch or lot can be identified.
	Medical device: indicates that the item is a medical device.
	Date of manufacture: indicates the date when the medical device was manufactured.
	Serial number: indicates the manufacturer's serial number so that a specific medical device can be identified.
	Consult instructions for use or consult electronic instructions for use: indicates the need for the user to consult the instructions for use.

	Keep dry: indicates a medical device that needs to be protected from moisture.
	Do not use if package is damaged and consult instructions for use: indicates that a medical device should not be used if the package has been damaged or opened and that the user should consult the instructions for use for additional information.
CE	CE label: indicates that the manufacturer or importer affirms the good's conformity with European health, safety, and environmental protection standards.
	WEEE: indicates that the product should not be discarded as unsorted waste but must be sent to separate collection facilities for recovery and recycling.
	Manufacturer: indicates the medical device manufacturer.

## Warranty

mo-vis bv warrants the product to be free from defects in material and workmanship for a period of 2 years under proper use, care and service. The dealer should never keep mo-vis products in stock for a period more than 6 months prior to delivery to the end-user. mo-vis' warranty will never exceed a period of 2 years and 6 months after shipment.

All warranties do not extend beyond the initial purchaser from an authorized mo-vis dealer or mo-vis itself.

### *Repair and replacement*

For warranty service, contact your dealer (or us if bought directly). In the event of a defect in material or workmanship, the dealer or customer must obtain a Return Merchandise Authorization (RMA) number from us. The product must be shipped to a service centre designated by mo-vis. mo-vis will repair or, at mo-vis' option, replace any product covered by the warranty.

### *Amendments*

No person is authorized to alter, extend or waive the warranties of mo-vis.

## ***Disclaimer and limitations of remedies***

The express warranties set forth in this agreement are in lieu of all other warranties of merchantability or fitness of purpose. In no event shall mo-vis be liable for any direct, indirect, incidental or consequential damages resulting from any defect in this product.

Warranty of parts subject to “normal wear and tear” (e.g. joystick handles, pads, ...) are not covered in the warranty except as it applies to defects in material or construction.

## ***Voiding of warranties***

The foregoing warranties are contingent upon the proper installation, use, maintenance and care of the product. The warranty will be void if the product has been installed or used improperly, or if it has been repaired or any part replaced by persons other than mo-vis or an authorized dealer. This product is considered as a non-serviceable part.

The addition of equipment or features that are not manufactured or recommended by mo-vis could affect the intended function of the mo-vis product and may invalidate the warranty.

## **Important information**



**CAUTION:** Incorrect use or installation may lead to risk of injury to the user and damage to the wheelchair or other property. In order to reduce these risks, you should carefully read this manual, paying particular attention to the safety instructions and warning texts.



**NOTICE:** Only install this product on a wheelchair where the wheelchair manufacturer allows the installation of third party parts.

## **Support, scrapping & recycling**

### ***Technical support***



**TROUBLE:** In case of technical problems:

- Contact your dealer
- If your dealer is not available or unknown, please contact mo-vis: [contact@mo-vis.com](mailto:contact@mo-vis.com) or +32 9 335 28 60.


Always state the product code and device serial number when contacting us. This ensures you are provided with the correct information.

### ***Spare parts and accessories***

Contact mo-vis or your dealer for more information about spare parts and accessories.

## Scrapping & recycling


The expected lifetime of this device is 5 years.


 **CAUTION:** For scrapping, adhere to your local waste legislation. Dispose of obsolete electronic parts responsibly in accordance with local recycling regulations.

## Intended use

The Twister:

- is an Input Control System for a device to be operated by a push button with a 3.5 mm jack connection ( e.g. power wheelchair or AAC device).
- only requires a light activation force of 30 gr.
- offers different mounting options for the different models.

 **CAUTION:** To make sure that the device maintains its functionality, regular cleaning is required. See [Cleaning on page 12](#) .

 **CAUTION:** This device is only intended to support auxiliary wheelchair functions.

## Features


The Twister:

- requires 30 gr of activation force.
- has an activation surface of 18 mm and 0.5 distance to activation.
- offers tactile and auditory (click) feedback.
- is available in 5 different models. See [Parts and accessories on page 8](#) .
- is easy to mount in different positions, using the mo-vis Q2M system.
- has a 3.5 mm mono jack plug.
- is available in 5 colours: yellow, green, red, blue and black.

## Operation

The Twister is a normally open switch. This means that it will activate when pressed down.

To activate your Twister, simply press anywhere on the cap of the device. You will feel and hear a clear and audible 'click'.

 **CAUTION:** Do not attempt to open or disassemble the device enclosure.





**NOTE:** Do not pull or wrap the cable around the device.


## Parts and accessories

### *Twister Basic*

The Twister Basic can be mounted on a plate of metal/plastic with two screws or can be integrated in a tray or other flat surface.

A Twister Basic package consists of the following parts:

PICTURE	DESCRIPTION
	1 piece: <ul style="list-style-type: none"><li>• P005-15/16/17/18/19 Twister Basic (300 mm) - black/red/yellow/blue/green</li><li>• P005-70/71/72/73/74 Twister Basic (1500 mm) - black/red/yellow/blue/green</li></ul>
	2 sets of 2 screws with diameter 2.2 mm/0.087 in: <ul style="list-style-type: none"><li>• Set with length of 6.5 mm/0.27 in for sheet thickness 0.5 mm/0.02 in - 2 mm/0.079 in.</li><li>• Set with length of 9.5 mm/0.37 in for sheet thickness 3 mm/0.19 in - 4.5 mm/0.18 in.</li></ul>

PICTURE	DESCRIPTION
	<p>Optionally, there is a M005-29 Twister Basic table mounting set available to mount the Twister Basic in a flat surface above/at/below surface level.</p>



### **Twister Button**

The Twister Basic is mounted in a plastic holder with a diameter of 30 mm/1.18 in to create a Twister Button.

The screw holes in the base plate are reinforced to attach the button firmly to a flat surface. The hole diameter is 2.5 mm/0.098 in and the distance between them is 8 mm/0.315 in.

The Twister Button is delivered with a set of easy-to-install and durable mounting pieces.

The Twister Button package consists of the following parts:

PICTURE	DESCRIPTION
	<p>1 P005-00 Twister Basic no screw</p>
	<p>M005-23 Twister Basic Mounting Set:</p> <ul style="list-style-type: none"> <li>• 1 Twister Basic Cone</li> <li>• 1 Self-adhesive rubber pad 28*2 mm/1.10*0.08 in</li> <li>• Set of 3M Dual Lock self-adhesive reclosable fasteners 25*2 mm/0.87*0.08 in</li> <li>• 2 sets of 2 screws with diameter 2.2 mm/0.087 in: <ul style="list-style-type: none"> <li>-Set with length of 6.5 mm/0.27 in for sheet thickness 0.5 mm/0.02 in - 2 mm/0.079 in.</li> <li>-Set with length of 9.5 mm/0.37 in for sheet thickness 3 mm/0.19 in - 4.5 mm/0.18 in.</li> </ul> </li> </ul>


This then becomes:

- P005-85/86/87/88/89 Twister Button (300 mm) - black/red/yellow/blue/green
- P005-75/76/77/78/79 Twister Button (1500 mm) - black/red/yellow/blue/green

### ***Twister on Bended Tube***

The Twister is attached to a mo-vis Q2M bended tube (length 10 cm/3.94 in - diameter 6 mm/0.236 in).


The Twister on Bended Tube can be mounted with the included Q2M clamps to attach to other rods or the power chair.

PICTURE	DESCRIPTION
	<p>1 piece:</p> <ul style="list-style-type: none"> <li>• P005-60/61/62/63/64 Twister on Bended Tube (300 mm) - black/red/yellow/blue/green</li> <li>• P005-50/51/52/53/54 Twister on Bended Tube (1500 mm) - black/red/yellow/blue/green</li> </ul>

### Satellite Twister

The Satellite Twister is attached to a mo-vis Q2M bended tube (length 10 cm/3.94 in - diameter 6 mm/0.236 in).


The Satellite Twister is equipped with an insert to mount it on a mo-vis All-round (Light) Joystick.

 **NOTE:** The Satellite Twister is only available with a short cable (300 mm).

PICTURE	DESCRIPTION
	<p>1 piece:</p> <ul style="list-style-type: none"> <li>• P005-40/41/42/43/44 Satellite Twister (300 mm) - black/red/yellow/blue/green</li> </ul>

### Twister Gooseneck

The Twister Gooseneck consists of a Twister Basic mounted on a flexible of 400 mm/15.75 in.

 **NOTE:** The Twister Gooseneck is only available with a long cable (1500 mm).


The Twister Gooseneck can easily be mounted with a Q2M 8 mm clamp (e.g. M004-14 Q2M Clamp D8, set = 10 pcs). The base of the Twister Gooseneck also fits perfectly in the M007-01 Q2M Universal Clamp and makes mounting on wheelchair tubes simple and easy (as seen in the picture below).


PICTURE	DESCRIPTION
	1 piece: <ul style="list-style-type: none"><li>• P018-20/21/24 Twister Gooseneck (1500 mm) - black/red/green</li></ul>


## Cleaning

Regular maintenance of the device is required. Before you do this, always disconnect the device.

- Gently remove dust and dirt with a damp cloth.
- Use only non-aggressive disinfectant cleaning agents.

 **CAUTION:** Do not immerse in water or use excessive amounts of liquid.

 **CAUTION:** Make sure the button stays clean so activation remains possible at all times.

 **CAUTION:** If the button is damaged, it must be replaced immediately to ensure correct functionality.

## Limited liability

*mo-vis accepts **no liability** for personal injury or damage to property that may arise from the failure of the user or other persons to follow the recommendations, warnings and instructions in this manual.*



**CAUTION:** This product should only be installed by a qualified service engineer.



**NOTE:** All mo-vis user manuals can be found at <http://www.mo-vis.com> where they can be consulted in PDF format.



**NOTE:** In case any serious incident occurs in relation to this device, this should be reported immediately to mo-vis and the competent authority of the Member State in which the user is established.



mo-vis bv . Biebuyckstraat 15D . 9850 Deinze - Belgium  
www.mo-vis.com . contact@mo-vis.com . +32 9 335 28 60

**Go to our website for more information on our products or share  
your experience with us via email.**

